

Adaptation and Validation of the Purcell method of IHNv detection



L.Louboutin, J.Cabon, M.Baud, L.Bigarré, T. Morin

Efficiency

Between 91 and 96% (slopes ranging from -3,4 to -3,5)

Repeatability

1 operator, 2 days						
	Nb of sessions	Operators	Nb of dilutions	Replicates/dilution		
	2	1	6	2		
Concentrations (PFU/mL)	1.10⁴	1.10⁴	1.10⁵	1.10⁴	1.10⁴	1.10⁵
06/05/2015	15,76	18,36	22,31	25,72	28,82	31,7
13/05/2015	15,78	18,51	22,48	25,99	29,43	32,45
Mean value	15,77	18,44	22,4	25,86	29,12	32,08
Standard deviation	0,08	0,15	0,11	0,16	0,36	0,72
CV%	0,51%	0,82%	0,51%	0,62%	1,22%	(2,25%)
CV < 3%						
Within operators						
	Nb of sessions	Operators	Nb of dilutions	Replicates/Dilution		
	2	1	6	2		
Concentrations (PFU/mL)	1.10⁴	1.10⁴	1.10⁵	1.10⁴	1.10⁴	1.10⁵
06/05/2015	14,3	18,77	22,37	26,26	29,57	33,31
11/06/2015	17,35	19,96	23,74	27,3	29,85	32,2
LL 06/05/2015	16,98	19,81	23,52	26,71	28,98	31,28
Mean value	16,88	19,52	23,34	26,76	29,47	32,20
Standard deviation	0,49	0,58	0,46	0,4889	0,40	0,93
CV%	2,93%	2,96%	1,95%	1,83%	1,36%	2,89%
CV < 6%						

Perspectives

① production of RNA Transcript to determine precisely LD_{PCR} (copies of RNA/reaction)

② Add an exogenous control to validate the whole process (extraction + amplification)

Choice : bacteriophage T4 (DNA) and MS2 (RNA) to spike samples before extraction

Diagnostic sensibility : 10 fold serial dilutions with and without control

③ Reproducibility : interested in participating a proficiency test for IHNv detection

④ To be accredited for IHNv detection by RT-PCR in september 2017
Transfer the method to the French accredited labs (7 labs)

Then deployment of the range of analysis (VHSV and KHV in 2018 ? IPNV in 2019 ?)
while maintaining accreditation in cellular virology ...

Protocol

QIAGEN Quantitect Probe Rt-PCR kit

Primer and probe concentrations → Purcell et al (900nM of each primers, 200nM of probe)

Reagents	Initial Concentration	Final Concentration	Volume for 1 reaction	Volume for 9 reactions	Programme
qPCR Probe master mix	20	4	0.250	2.25	1. 30 min 50°
pVP27	20	900	1.13	10.13	2. 15 min 95°
pVP23	20	900	1.13	10.13	3. 15 sec 94°
H2O			0.250	2.25	4. 60 sec 60° ***
qRT mix			0.250	2.25	call endpoint data FAM
			4.75	42.75	5. goto 3. 39x
Volume total (μ)			20.00	180.00	
Added	5	μ ARN to	20	μ MasterMix	

Tests of ruggedness

➢ Slight variations for :

- Primers and probe concentrations (+/- 20%)

- PCR program

➢ Two-step versus one-step

➢ 5 ten-fold dilutions (supernatant of N61, internal French reference strain)

No significant CT difference with +/- 20% [primers] or [probe]

Analytical specificity

	Sample status			100% positive supernatant detected
	Status delivered by the method	Positive	Negative	
40 supernatants		34 true positive	0 false positive	
	Negative	0 False Negative	6 True Negative	

Analytical sensitivity (limit of detection)

	Nb of séances	Operators	Nb of dilutions	Replicates/dilution
LOD	3	1	6	8
DNA plasmid	3	1	6	2
N61 RNA				

- 25 copies / reaction using plasmid pPV22 (Ct=33,4)
- 10 PFU / reaction using N61 IHNv strain (2.10³ PFU/mL); Ct=32,6

Diagnostic specificity : Dsp = 100%

	Sample status		
	Status delivered by the method	Positive	Negative
		0 true positive	0 false positive
	Negative	0 False Negative	2 True Negative

Diagnostic sensitivity

2 methods :

- organ homogenate spiked with plasmidic DNA = 10 cp / reaction (Ct=34,5)

- organs collected after experimental infections ; serial dilutions (Ct=35,2)

Quite difficult to find Dse (random signal with late Ct)